



EU DECLARATION OF CONFORMITY

Product Identification:

Fluke 51II, 52II, 53II, 54II, 53IIB, 54IIB - Thermometer

First declared: 2007

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO9001:2008 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2011/65/EU, RoHS

Directive 2014/30/EU, Electromagnetic Compatibility (EMC)

Directive 2014/35/EU, Low Voltage (LVD)

Standards Used:

EN 50581: 2012 Restriction of the use of hazardous substances

EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General

EN 61326-2-2: 2013 Particular EMC requirements for portable test, measuring and monitoring equipment

EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General

Special Conditions:

None

Manufacturer

Fluke
6920 Seaway Boulevard M/S 143F
Everett, WA 98203; USA

Fluke Europe B.V.
Science Park Eindhoven 5110, 5692 EC Son
The Netherlands

Thomas M. Smith P.E.
Product Compliance Manager, Fluke Corporation
DC20161004084636 - This declaration supersedes all previous declarations for this product.

Alma Miklyciute
EMEA Trade Compliance Manager

04 October 2016