



EU DECLARATION OF CONFORMITY

Product Identification:

Fluke 353, 355 - AC/DC Clamp Meter

First declared: 2007

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO9001:2008 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2014/30/EU, Electromagnetic Compatibility (EMC)

Directive 2014/35/EU, Low Voltage (LVD)

Standards Used:

EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General
EN 61326-2-2: 2013 Particular EMC requirements for portable test, measuring and monitoring equipment
EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General
EN 61010-2-032: 2002 Particular requirements for hand-held and hand manipulated current sensors

Special Conditions:

None

Manufacturer

Fluke
6920 Seaway Boulevard M/S 143F
Everett, WA 98203; USA

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC20160812092743 - This declaration supersedes all previous declarations for this product.

12 August 2016