


Supplement

Title: 1AC II Inst. Sht. Supplement Issue: 2
Print Date: October 2009 Issue Date: 11/17
Part Number: Multiple Page Count: 1
Revision/Date:

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 64826, 64873

Add the following to the **Symbols** table:

	Conforms to relevant South Korean EMC Standards.
---	--

Following the **Symbols** table, delete the **Safety Compliance** and the **EMC Compliance** section and add:

Safety.....IEC 61010-1


Electromagnetic Environment....IEC 61326-1: Portable

Electromagnetic Compatibility...Applies to use in Korea only.
Class A Equipment (Industrial
Broadcasting &
Communication Equipment) ^[1]

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.

Under **Operating Instructions**, replace the existing text with:

Turning the Tester On

Momentarily push . Listen for a double beep to confirm activation. A continual VoltBeat flash indicates the Tester is active.

Change #2, 511

In the Symbols table, replace the 6th row with:

	Conforms to Australian standards.
---	-----------------------------------