

Instrument Security Procedures

Model:

Fluke 27II and 28II

Product Name:

Digital Multimeters

Instrument Description:

Fluke 2x II Digital Multimeters

Memory Description:

Fluke 27II and 28II have one memory device: U2, micro controller, TI MSP430F449, 60kB Flash, 2048B RAM, contains operating code for the product and calibration constants.

Memory Cleaning Instructions:

The user controlled storage operation other than calibration constants is the storage of Relative, Min Max and Average. This memory is cleared when the unit is powered off.