

# Manual Supplement

Manual Title:	233 Calibration	Supplement Issue:	<b>2</b>
Part Number:	Web-only	Issue Date:	4/14
Print Date:	October 2009	Page Count:	1
Revision/Date:			

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1, 63818, 65340

On page 3, add the following to the Symbols table:

☐	Conforms to relevant South Korean EMC Standards.
C	Conforms to CAN/CSA-C22.2 No. 61010-1, second edition, including Amendment 1.
CAT II	Measurement Category II is applicable to test and measuring circuits connected directly to utilization points (socket outlets and similar points) of the low-voltage MAINS installation.
CAT III	Measurement Category III is applicable to test and measuring circuits connected to the distribution part of the building's low-voltage MAINS installation.
CAT IV	Measurement Category IV is applicable to test and measuring circuits connected at the source of the building's low-voltage MAINS installation.

On page 3, under **General Specifications**, remove **Electromagnetic Compatibility, Safety Compliance** and **Certifications** and add:

**Electromagnetic Compatibility** ..... In an RF field of 3 V/m, accuracy = specified accuracy except in temperature:  
 specified accuracy  $\pm 5$  °C (9 °F)  
 FCC ID T68-F233  
 IC: 6627A-F233

Applies to use in Korea only ..... Class A Equipment (Industrial Broadcasting & Communication Equipment) <sup>[1]</sup>

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.

**Safety** ..... IEC 61010-1: Pollution Degree 2  
 IEC 61010-2-030: 600 V CAT IV / 1000 V CAT III

## Change #2

On page 13, Table 3, replace step 18 with:

18	$\sim$ Hz <b>V</b> AC Volts + Hz	6.000 V	2 V, 50 kHz <sup>[1]</sup>	49.97 to 50.03
----	--	---------	----------------------------	----------------